

On-site calibration and verification of mass flowmeter for aerosol samplers

Lise Hégron, CT2M

lhegron@ct2m.fr



LE PARTENAIRE À VOTRE MESURE

www.ct2m.fr

CT2M in figures

1993 : creation

11 people

660 people trained in 2022

3074 masses calibrated in 2022

5 CIL organized in 2022

✉ Centre des Creusets
Route de Lançon
13250 SAINT CHAMAS

📞 04 90 50 90 14

💻 www.ct2m.fr



Training / Consulting / Audit@

« Supporting companies and laboratories in their Quality and Metrology projects »

Contact: David BENHAMOU

dbenhamou@ct2m.fr

+33 (0)6.78.00.10.26



« Actions de formation »

Proficiency Testing provider (PT)

« Organization of PT in accordance with ISO 17043 and with COFRAC accreditation for mass calibration from 1 mg to 20 kg »

Contact: Boris GEYNET

bgeynet@ct2m.fr

+33 (0)6.83.94.60.87



Accreditation N° 1-7127

Scope available on www.cofrac.fr

Mass calibration until E2 class

« Calibration and verification of masses from 1 mg to 5 tons with COFRAC accreditation »

Contact: Anaïs LAMOUR

alamour@ct2m.fr

+33 (0)4.90.50.90.14




Accreditation N° 2-1292

Scope available on www.cofrac.fr

Context : Measurement of radioactivity in the environment


Measurement : global alpha / beta radioactivity index or gamma emitters in environmental samplers

Unit: **Bq/Nm³** (Normal Meter Cubed)



Metrological traceability of the volume must be guaranteed

Mass flowmeter calibration (on-site)



Problems encountered by the company

Calibration laboratories of mass flowmeter are accredited “in laboratory” and not “on site” (aerosol sampler cannot be moved).



Studie

Internal calibration of mass flowmeter



Requirements of GEN REF 10 (COFRAC document based on ILAC P/10*)

Appropriate evidence for the technical competence to perform internal calibration and metrological traceability should include ISO 17025 requirements.

**ILAC Policy on Metrological Traceability of Measurement Results*

Requirements of GEN REF 10 (COFRAC document based on ILAC P/10*)

Appropriate evidence for the technical competence to perform internal calibration and metrological traceability should include ISO 17025 requirements.

- 1) Metrological criteria
- 2) Calibration method of mass flowmeter (+ environmental conditions)
- 3) Calibration results
- 4) Calibration uncertainty
- 5) Statement of compliance

**ILAC Policy on Metrological Traceability of Measurement Results*

1) Metrological criteria:

Requirements of the NF M 60-760 standard:*

Requirement on:	Criteria	Value
Volume measurement of air sample	Relative uncertainty	< 5% (k=1)
Proper functioning of the flow control device	Deviation from the set point	±10%

Company criteria : ±2%


*NF M 60-760 (2017): Nuclear energy - Measurement of radioactivity in the environment - Air - Sampling of aerosols for measurement of radioactivity in the environment

2) Calibration method of mass flowmeter:

3 tests are carried out with a **mass flowmeter standard**:

- Flow calibration (in Nm^3/h)
- Volume control (in Nm^3)
- Control of flow regulation (in Nm^3/h)

→ *Time quantity is not calibrated because the internal clock error is insignificant.*



2) Calibration method of mass flowmeter:

Measuring range :

- based on $\pm 10\%$ deviation criteria
- Set point : $6,25 \text{ Nm}^3/\text{h}$



5,6 to 6,9 Nm^3/h

Calibration point:

- Flow calibration (in Nm^3/h)
- Control of flow regulation (in Nm^3/h)



5,6 Nm^3/h

6,25 Nm^3/h

6,9 Nm^3/h



Modification of the set point
locked in the software

2) Calibration method of mass flowmeter + environmental conditions


Environmental conditions:

- Air temperature between 15°C and 25°C
- Absence of strong winds
- Avoid humid periods






Operating mode:

- Flow calibration (in Nm³/h) :
- Control of the flow regulation (in Nm³/h)
- Volume control (in Nm³)




3) Calibration results:

- Flow calibration (in Nm³/h)  **Measurement bias « b_{flow} »**
- Volume control (in Nm³)  **Measurement bias « b_{vol} »**
- Control of flow regulation (in Nm³/h)  **Deviation from the set point « d_{regul} »**

4) Calibration uncertainty:

→ GUM methodology

- Flow calibration : $U(b_{\text{flow}})$ in Nm^3/h
- Volume control : $U(b_{\text{vol}})$ in %



Results example:

Set point : $6,25 \text{ Nm}^3/\text{h}$

- $U(b_{\text{flow}}) = 0,13 \text{ Nm}^3/\text{h}$
- $U(b_{\text{vol}}) = 2,1 \%$

5) Statement of compliance:

→ In relation to Maximum Permissible Error (MPE)

- Flow calibration : $|b_{\text{flow}}| + U(b_{\text{flow}}) \leq \text{MPE}_{f,v}$
- Volume control : $|b_{\text{vol}}| + U(b_{\text{vol}}) \leq \text{MPE}_{f,v}$
- Control of the flow regulation : $|d_{\text{regul}}| \leq \text{MPE}_d$

5) Statement of compliance:

→ Determination of Maximum Permissible Error on flow and volume ($MPE_{f,v}$)

- Criteria :

Requirement on:	Criteria	Value
Volume measurement of air sample	Uncertainty	< 5% (k=1)

- Evaluation of air sample volume uncertainty:

$$u_{rel}(\text{massflowmeter}) = \frac{MPE_{f,v}}{\sqrt{3}}$$

$$u_{rel}(\text{flow variability}) = \frac{MPE_d}{\sqrt{3}}$$



$$u_{rel}(\text{volume of air sample}) < 5\% (k=1) \text{ with } \begin{cases} MPE_{f,v} = 8\% \\ MPE_d = 2\% \end{cases}$$

CONCLUSION

- **Development of an internal method for on-site mass flowmeter calibration :**
 - ✓ Validation of calibration method
 - ✓ Evaluation of measurement calibration uncertainty
 - ✓ Documentation and records for metrological traceability of measurement results
 - ✓ Documentation and records for competence of personnel
 - ✓ Documentation for facilities and environmental conditions

- **Areas of improvement :**
 - ✓ Software access
 - ✓ Uncertainty evaluation
 - ✓ Environmental condition control